



New laboratory measurements of the temperature-dependence of heterogeneous removal of N₂O₅ by sulfate aerosols

P. Griffiths (1), P. Cassanelli (1) and R. A. Cox (1)

(1) Centre for Atmospheric Science, Chemistry Department, Cambridge University, Lensfield Road, Cambridge, CB2 1EW, UK. (ptg21@cam.ac.uk / +44-1223-336362)

Nitrogen oxides play a central role in the chemistry of the atmosphere. Heterogeneous removal of N₂O₅ can be a major loss route for NO_x with recent modelling work by Tie et al (2003) suggesting that, at high latitudes, N₂O₅ hydrolysis can reduce NO_x levels by as much as 90 %.

The majority of laboratory studies of the heterogeneous uptake of N₂O₅ by aerosol particles have been performed at room temperature. At present, the role of temperature is the largest source of uncertainty in global loss rates. We are undertaking a series of laboratory measurements of the rate of removal of N₂O₅ by aerosol particles using a new apparatus capable of reaching temperatures relevant to the polar troposphere. These complement our earlier measurements on sulfates (Hallquist et al., 2003) and sulfates mixed with high molecular weight organics (Badger et al., 2006).

The instrument has been developed in conjunction with Tapcon, Austria, and comprises a temperature-controlled atmospheric pressure aerosol flow tube and differential mobility analyser. The operating range of the instrument is 240-320 K, with control of relative humidity over the range 0-80%.

We present a description of the apparatus and new measurements of the uptake coefficient, γ , for loss of N₂O₅ made using the new apparatus.